

ORDERING INFORMATION >>>

DIAMONDBACK 360® CORONARY ORBITAL ATHERECTOMY SYSTEM



**CLASSIC CROWN
MODEL NUMBERS**

CROWN SIZE

SHAFT LENGTH

QUANTITY

DBEC-125

1.25 mm

135 cm

1 each



SIP-3000

Saline Infusion Pump

1 each

VIPERWIRE Advance® CORONARY GUIDE WIRE



**MODEL
NUMBER**

DESCRIPTION

**WIRE
LENGTH**

QUANTITY

GWC-12325LG-FLP

0.012"/0.014" tip

325 cm

Box of 5

Only for use with CSI Diamondback 360 Coronary Orbital Atherectomy System.

VIPERSLIDE® LUBRICANT



**MODEL
NUMBER**

QUANTITY

VPR-SLD2

Box of 10 Bags

Indications: The Diamondback 360® Coronary Orbital Atherectomy System (OAS) is a percutaneous orbital atherectomy system indicated to facilitate stent delivery in patients with coronary artery disease (CAD) who are acceptable candidates for PTCA or stenting due to de novo, severely calcified coronary artery lesions.

Contraindications: The OAS is contraindicated when the ViperWire™ guide wire cannot pass across the coronary lesion or the target lesion is within a bypass graft or stent. The OAS is contraindicated when the patient is not an appropriate candidate for bypass surgery, angioplasty, or atherectomy therapy, or has angiographic evidence of thrombus, or has only one open vessel, or has angiographic evidence of significant dissection at the treatment site and for women who are pregnant or children.

Warnings/Precautions: Performing treatment in excessively tortuous vessels or bifurcations may result in vessel damage; The OAS was only evaluated in severely calcified lesions. A temporary pacing lead may be necessary when treating lesions in the right coronary and circumflex arteries; On-site surgical back-up should be included as a clinical consideration; Use in patients with an ejection fraction (EF) of less than 25% has not been evaluated.

See the instructions for use before performing Diamondback 360 coronary orbital atherectomy procedures for detailed information regarding the procedure, indications, contraindications, warnings, precautions, and potential adverse events. For further information call CSI at 1-877-274-0901 and/or consult CSI's website at www.csi360.com.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician



651 Campus Drive
St. Paul, MN 55112

T: 651.259.1600
877.274.0901

F: 612.677.3355
www.csi360.com

For more information, please contact your local CSI representative or call customer service at 1.877.274.0901.

CSI, Diamondback 360, ViperWire Advance and ViperSlide are registered trademarks of Cardiovascular Systems, Inc. ©2013 Cardiovascular Systems, Inc.